




# Overall Height of Liko Slings

English  
7EN160828 Rev. 4



## Overall Sling Height

The distance from the sling bar to the lowest point of the seat.  
Note that this is an approximate measure that may vary.

Product Name	Size	Overall Sling Height	
		cm	inches
UniversalSling, Mod. 00, 000	S	75	29.5
	MS (Medium-Slim)	85	33.5
	M	85	33.5
	L	95	37.4
	XL	95	37.4
Universal HighLevel Sling, Mod. 02, 002	M	75	29.5
	L	85	33.5
OriginalSling/TeddySling Original, Mod. 10/11	XS	75	29.5
	S	80	31.5
	MS (Medium-Slim)	75	29.5
	M	80	31.5
	L	90	35.4
	XL	90	35.4
Original HighBack Sling, Mod. 20/21, 200/210	XS	80	31.5
	S	90	35.4
	MS (Medium-Slim)	95	37.4
	M	95	37.4
	L	105	41.3
	XL	105	41.3
	XXL	115	45.3
SilhouetteSling, Mod. 22	XS	75	29.5
	S	80	31.5
	M	105	41.3
	L	105	41.3
Solo HighBack Sling, Mod. 25	MS (Medium-Slim)	85	33.5
	M	90	35.4
	L	100	39.4
	XL	105	41.3
Soft Original HighBack Sling/TeddySling Soft Original, Mod. 26	XS	80	31.5
	S	90	35.4
	MS (Medium-Slim)	95	37.4
	M	95	37.4
	L	105	41.3
	XL	105	41.3
	XXL	110	43.3



Product Name	Size	Overall Sling Height	
		cm	inches
ComfortSling Plus, Mod. 300, 310	M	85	33.5
	L	90	35.4
	XL	95	37.4
ComfortSling Plus High, Mod. 350, 360	M	90	35.4
	L	90	35.4
	XL	100	39.4
HygieneSling/Teddy Hygiene SafetySling , Mod. 40	XS	65	25.6
	S	75	29.5
	M	90	35.4
	L	95	37.4
	XL	95	37.4
HygieneSling with Belt/Teddy HygieneSling , Mod. 41	XS	65	25.6
	S	85	33.5
	M	85	33.5
	L	90	35.4
	XL	110	43.3
HygieneSling/Teddy Hygiene SafetySling , Mod. 45, 46	XS	65	25.6
	S	80	31.5
	M	95	37.4
	L	100	39.4
	XL	95	37.4
HygieneVest /Teddy HygieneVest, Mod. 50, 55	XXS	70	27.6
	XS	80	31.5
	S	95	37.4
	M	100	39.4
	L	105	41.3
	XL	105	41.3
AmputeeSling, Mod. 70	S	80	31.5
	M	85	33.5
	L	90	35.4
Amputee HighBack Sling, Mod. 75	S	90	35.4
	M	95	37.4
	L	100	39.4
	XL	110	43.3

All measurements are taken with the outermost strap loops connected to the sling bar. The size of the patient (how much of the sling the patient takes up) determines the overall height of the sling. All the overall height measurements for HygieneSlings and HygieneVests are taken on patients with good muscle tone; poor muscle tone may increase the overall sling height.



A Company

[www.liko.com](http://www.liko.com)

**Manufacturer:**

Liko AB  
Nedre vägen 100  
SE-975 92 Luleå  
Sweden